

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Illiana Instrumentation
1831 Govert Drive
Schererville, IN 46375-

Date received: 10/14/20
Calibration date: 10/14/20
Recalibration due: 10/31/21

Purchase order:
Location of calibration: Lab

Item number: 1245
Tag number:
Manufacturer: Ametek
Model num: 250SE
Description: 1245 Ametek
Serial number: 010115-01093
Accuracy: +/- 1.6 deg F.
Item notes:

Item location: Lab/Mobile
Temp: 74.0° F. Humidity: 30 Sensitivity: OK
Condition as found: In tolerance
Condition as left: In tolerance
Intentional offset as found: 0.0000
Intentional offset as left: 0.0000
Limitations: None

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the laboratory may cause the calibration of the above item to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for review upon request and approval of our Quality Manager. The uncertainties represent an expanded uncertainty expressed at @95% confidence level using a coverage factor of k=2. The date this report is signed constitutes the issue date. Pass/Fail criteria does not take into account measurement uncertainty.

Tech: Paul D Grolla Date: 10/14/20 Approved by: Laura Grolla Date: 10/14/20
Paul D Grolla - ISA Certified Controls Systems Tech. Laura Grolla - Deputy Quality Manager

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
122.0000	122.9000	122.9000	123.6000	120.4000	Deg F.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.1200 Deg F.

Procedure: WI-516

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
212.0000	212.4700	212.4700	213.6000	210.4000	Deg F.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.1200 Deg F.

Procedure: WI-516

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
302.0000	301.9000	301.9000	303.6000	300.4000	Deg F.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.1200 Deg F.

Procedure: WI-516

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
392.0000	391.4000	391.4000	393.6000	390.4000	Deg F.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.1200 Deg F.

Procedure: WI-516



ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
482.0000	480.8000	480.8000	483.6000	480.4000	Deg F.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.1200 Deg F.

Procedure: WI-516

Customer approval: _____ Date: _____

Total pages for Certificate 999999-01245-5464: 2