



# ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 Phone (219)227-8788 Fax (219)515-6161

## CERTIFICATE OF CALIBRATION

CUSTOMER:	MISCELLANEOUS DETAILS:
Illiana Instrumentation Service 1831 Govert Drive Schererville, IN 46375	Date Received 3/10/22 Certification Date: 3/10/22 Recalibration Date: 6/10/22 Cal. Number: 1336-031022 P.O. Number: Location of Calibration: Lab Detailed Results Attached: YES Procedure Used: Fluke Procedure

EQUIPMENT CALIBRATED	
MANUFACTURER:	Fluke
MODEL:	725
SERIAL NUMBER:	9531036
ITEM NUMBER:	1336
DESCRIPTION:	Calibrator
CONDITION AS FOUND:	In tolerance
LIMITATIONS:	None

STANDARDS USED
Item 1546 Fluke 525B; Item 1205 Hewlett Packard 34401A

TEST CONDITIONS	
TEMPERATURE	73 Deg F.
HUMIDITY	38% rH

CERTIFIED BY: Paul Droll TITLE: ISA CCST III DATE 3/10/22

APPROVED BY: Jana Droll TITLE: Asst. Quality Mgr. DATE 3/10/22

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for examination upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2. The date this report is signed constitutes the issue date. Pass/Fail criteria does not take into account measurement uncertainty.

Item	1336 Fluke 725								
Accuracy:	Varies By Range								
Date:	3/10/2022								
Intentional Offset as Found	None								
Intentional Offset as Left	None								
Limitations	None								
Input range	Eng. Units	Cal Pt	Upper	lower	Initial	Final	Sensitivity	Unc.	
							Check		
Meas VDC Upper Display	VDC	0	0.002	-0.002	0	0	OK	0.000005	
		15	15.005	14.995	14.998	14.998	OK	0.000005	
		30	30.008	29.992	29.997	29.997	OK	0.000005	
Meas mVDC on lower display	mDVC	0	0.02	-0.02	0	0	OK	0.0032	
		45	45.03	44.97	45	45	OK	0.0032	
		90	90.04	89.96	90	90	OK	0.0032	
Lower display voltage measure	VDC	0	0.002	-0.002	0	0	OK	0.000005	
		10	10.004	9.996	9.998	9.998	OK	0.000005	
		20	20.006	19.994	19.998	19.998	OK	0.000005	
Meas mADC 30 mADC upper display	mADC	4	4.003	3.997	4	4	OK	0.0016	
		12	12.005	11.995	11.999	11.999	OK	0.0016	
		24	24.007	23.993	23.998	23.998	OK	0.0057	
Meas mADC 30 mADC lower display	mADC	4	4.003	3.997	4	4	OK	0.0016	
		12	12.005	11.995	11.999	11.999	OK	0.0016	
		24	24.007	23.993	23.999	23.999	OK	0.0057	
Frequency source lower display	kHz	10	10.025	9.975	9.999	9.999	OK	0.0033	
Lower display 4 wire resistance measure	Ohms	15	15.1	14.9	15.01	15.01	OK	0.025	
		350	350.1	349.9	350.02	350.02	OK	0.025	
		500	500.5	499.5	500	500	OK	0.049	
Lower display 3 wire RTD measure	Ohms	350	350.2	349.8	350.03	350.03	OK	0.049	
Measure K	C	-180	-179.4	-180.6	-180.3	-180.3	OK	0.17	
Measure K	C	0	0.6	-0.6	-0.1	-0.1	OK	0.17	
Measure K	C	400	400.6	399.4	399.8	399.8	OK	0.17	
Measure K	C	800	800.8	799.2	799.8	799.8	OK	0.17	
Measure K	C	1000	1001	999	999.8	999.8	OK	0.17	
Measure K	C	1300	1301.3	1298.7	1299.8	1299.8	OK	0.17	
Simulate K	C	-180	-179.4	-180.6	-179.7	-179.7	OK	0.16	
Simulate K	C	0	0.6	-0.6	0.2	0.2	OK	0.16	
Simulate K	C	400	400.6	399.4	400.2	400.2	OK	0.16	
Simulate K	C	800	800.8	799.2	800.1	800.1	OK	0.16	
Simulate K	C	1000	1001	999	1000.1	1000.1	OK	0.16	
Simulate K	C	1300	1301.3	1298.7	1300	1300	OK	0.16	
Measure J	C	-200	-199.4	-200.6	-200.3	-200.3	OK	0.15	
Measure J	C	0	0.6	-0.6	-0.1	-0.1	OK	0.15	
Measure J	C	300	300.6	299.4	299.9	299.9	OK	0.15	
Measure J	C	600	600.6	599.4	599.8	599.8	OK	0.15	
Measure J	C	900	900.9	899.1	899.9	899.9	OK	0.15	
Measure J	C	1200	1201.2	1198.8	1199.8	1199.8	OK	0.15	
Simulate J	C	-200	-199.4	-200.6	-199.6	-199.6	OK	0.13	
Simulate J	C	0	0.6	-0.6	0.2	0.2	OK	0.13	
Simulate J	C	300	300.6	299.4	300.2	300.2	OK	0.13	
Simulate J	C	600	600.6	599.4	600.2	600.2	OK	0.13	
Simulate J	C	900	900.9	899.1	900.1	900.1	OK	0.13	
Simulate J	C	1200	1201.2	1198.8	1200.1	1200.1	OK	0.13	

Input range	Eng. Units	Cal Pt	Upper	lower	Initial	Final	Sensitivity	Unc.
							Check	
Measure S	C	1000	1001	999	999	999	OK	0.67
Measure S	C	1150	1151.5	1148.5	1149	1149	OK	0.58
Measure S	C	1300	1301.3	1298.7	1299	1299	OK	0.58
Measure S	C	1450	1451.5	1448.5	1449	1449	OK	0.58
Measure S	C	1600	1601.6	1598.4	1599	1599	OK	0.58
Measure S	C	1750	1751.75	1748.25	1749	1749	OK	0.58
Simulate S	C	1000	1001	999	1000	1000	OK	0.67
Simulate S	C	1150	1151.5	1148.5	1149.9	1149.9	OK	0.58
Simulate S	C	1300	1301.3	1298.7	1300	1300	OK	0.58
Simulate S	C	1450	1451.5	1448.5	1449.9	1449.9	OK	0.58
Simulate S	C	1600	1601.6	1598.4	1600	1600	OK	0.58
Simulate S	C	1750	1751.75	1748.25	1749.9	1749.9	OK	0.58
Measure N	C	0	0.6	-0.6	-0.4	-0.4	OK	0.2
Measure N	C	200	200.6	199.4	199.8	199.8	OK	0.2
Measure N	C	500	500.6	499.4	499.7	499.7	OK	0.2
Measure N	C	800	800.8	799.2	799.7	799.7	OK	0.2
Measure N	C	1100	1101.1	1098.9	1099.7	1099.7	OK	0.2
Measure N	C	1300	1301.3	1298.7	1299.7	1299.7	OK	0.2
Source N	C	0	0.6	-0.6	0.3	0.3	OK	0.18
Source N	C	200	200.6	199.4	200.3	200.3	OK	0.18
Source N	C	500	500.6	499.4	500.3	500.3	OK	0.18
Source N	C	800	800.8	799.2	800.2	800.2	OK	0.18
Source N	C	1100	1101.1	1098.9	1100.2	1100.2	OK	0.18
Source N	C	1300	1301.3	1298.7	1300.2	1300.2	OK	0.18
Measure T	C	-200	-199.4	-200.6	-199.8	-199.8	OK	0.74
Measure T	C	0	0.6	-0.6	0	0	OK	0.35
Measure T	C	100	100.6	99.4	100	100	OK	0.35
Measure T	C	200	200.6	199.4	200	200	OK	0.35
Measure T	C	300	300.6	299.4	300	300	OK	0.35
Measure T	C	400	400.6	399.6	400	400	OK	0.35
Source T	C	-200	-199.4	-200.6	-199.7	-199.7	OK	0.73
Source T	C	0	0.6	-0.6	0.1	0.1	OK	0.33
Source T	C	100	100.6	99.4	100.1	100.1	OK	0.33
Source T	C	200	200.6	199.4	200	200	OK	0.33
Source T	C	300	300.6	299.4	300	300	OK	0.33
Source T	C	400	400.6	399.6	400	400	OK	0.33
Source mADC lower display	mADC	4	4.0028	3.9972	4.0004	4.0004	OK	0.01
		12	12.0044	11.9956	12.001	12.001	OK	0.01
		24	24.0068	23.9932	24.001	24.001	OK	0.012
Source mVDC lower display	mVDC	0	0.02	-0.02	0	0	OK	0.0045
		45	45.03	44.97	44.993	44.993	OK	0.01
		100	100.04	99.96	99.986	99.986	OK	0.01
Source VDC lower display	VDC	0	0.002	-0.002	0.00003	0.00003	OK	0.0000045
		5	5.003	4.997	4.9980	4.9980	OK	0.00053
		10	10.004	9.996	9.9960	9.9960	OK	0.0058
Lower display ohm source	Ohms	15	15.1	14.9	15.01	15.01	OK	0.0018
		360	360.1	359.9	360.02	360.02	OK	0.016
		500	500.5	499.5	500.06	500.06	OK	0.016
		1500	1500.5	1499.5	1500.5	1500.5	OK	0.016
		3200	3201	3199	3200.5	3200.5	OK	0.016