

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
1000.0000	999.7000	999.7000	1001.0000	999.0000	Deg F.	Type K	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 1.1000 Deg F.

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
2192.0000	2192.0000	2192.0000	2194.0000	2190.0000	Deg F.	Type K	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 1.1000 Deg F.

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	99.5000	99.5000	101.0000	99.0000	Deg F.	Type N	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 0.3000 Deg F.

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
450.0000	449.5000	449.5000	451.0000	449.0000	Deg F.	Type N	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 0.3000 Deg F.

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
780.0000	779.5000	779.5000	781.0000	779.0000	Deg F.	Type N	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 0.3000 Deg F.

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
1110.0000	1110.0000	1110.0000	1111.0000	1109.0000	Deg F.	Type N	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 1.1000 Deg F.

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
1440.0000	1439.0000	1439.0000	1441.0000	1439.0000	Deg F.	Type N	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 1.1000 Deg F.

Procedure: WI-508

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
1800.0000	1799.0000	1799.0000	1801.0000	1799.0000	Deg F.	Type N	Yes

Standard(s): 1546 Fluke 525B

Adjustments:

Problems:

Uncertainty: 1.1000 Deg F.

Procedure: WI-508

Customer approval: _____ Date: _____

Total pages for Certificate 999999-01388-5472: 3