

Certificate of Calibration



Illiana Instrumentation Service LLC
1831 Govert Drive
Scherverville, IN 46375



Cert #: 4102 Rev. 1 Item # 1468 Certificate Date: 9/1/2022

CUSTOMER

ILLIANA INSTRUMENTATION SERVICE LLC
1831 GOVERT DRIVE
SCHERERVILLE, IN 46375

Purchase Order:

EQUIPMENT INFORMATION

Item # 1468 Calibration Result: PASS
Manufacturer FLUKE
Model Number 725
Serial Number 1100083
Equip. Location MOBILE
Description PROCESS CALIBRATOR

CALIBRATION INFORMATION

Date received: 9/1/2022 Cal Date: 9/1/2022 Cal. Interval 3 MONTHS Cal. Due Date: 12/1/2022
Technician: CODY NIMON Temperature: 73 F Humidity 48 Cal location LAB
Tag

Department:

Notes:

Problems: None

Adjustments None

Sensitivity OKAY

Initial modification offset 0

Final modification offset 0

Note: If the modification offset is not zero, readings reflect zero.

Limitations NONE

Accuracy VARIES BY POINT

Quality and traceability statements

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for examination upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2. The certificate date on this report constitutes the issue date. Pass/Fail criteria does not take into account measurement uncertainty.

Calibrated By:

Approved By:

Laura Grolla

Customer Approval _____

Cert #: 4102 Rev. 1

Item # 1468

Certificate Date: 9/1/2022

Calibration Data

Values found OOT and adjusted are indicated by "A" and items left OOT are indicated by "F"

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT
MEASURE	VDC UPPER	0.0000	-0.0020	0.0020	0.000	0.000	VDC	0.000005	
MEASURE	VDC UPPER	15.0000	14.9950	15.0050	15.000	15.000	VDC	0.0051	
MEASURE	VDC UPPER	30.0000	29.9920	30.0080	30.000	30.000	VDC	0.0051	
		0	0	0					
MEASURE	mVDC LOWER	0.0000	-0.0200	0.0200	0.000	0.000	mVDC	0.0033	
MEASURE	mVDC LOWER	45.0000	44.9700	45.0300	44.990	44.990	mVDC	0.0033	
MEASURE	mVDC LOWER	90.0000	89.9600	90.0400	90.000	90.000	mVDC	0.0033	
		0	0	0					
MEASURE	VDC LOWER	0.0000	-0.0020	0.0020	0.000	0.000	VDC	0.000005	
MEASURE	VDC LOWER	10.0000	9.9960	10.0040	9.998	9.998	VDC	0.0051	
MEASURE	VDC LOWER	20.0000	19.9940	20.0060	19.996	19.996	VDC	0.0051	
		0	0	0					
MEASURE	mADC UPPER	4.0000	3.9970	4.0030	4.000	4.000	maDC	0.0016	
MEASURE	mADC UPPER	12.0000	11.9950	12.0050	12.001	12.001	maDC	0.0016	
MEASURE	mADC UPPER	24.0000	23.9930	24.0070	24.002	24.002	maDC	0.0057	
		0.00	0.00	0.00					
MEASURE	mADC LOWER	4.0000	3.9970	4.0030	4.000	4.000	maDC	0.0016	
MEASURE	mADC LOWER	12.0000	11.9950	12.0050	12.002	12.002	maDC	0.0016	
MEASURE	mADC LOWER	24.0000	23.9930	24.0070	24.002	24.002	maDC	0.0057	
		0.0	0.0	0.0					
MEASURE	FREQ. LOWER	10.0000	9.9800	10.0200	10.000	10.000	kHz	0.0025	
		0.0	0.0	0.0					
SOURCE	FREQ. LOWER	10.00000	9.97500	10.02500	9.99984	9.99984	kHz	0.0033	
		0.0	0.0	0.0					
MEASURE	OHMS LOWER	15.00	14.90	15.10	14.98	14.98	Ohm	0.025	
MEASURE	OHMS LOWER	350.00	349.90	350.10	350.01	350.01	Ohm	0.025	
MEASURE	OHMS LOWER	500.00	499.50	500.50	500.0	500.0	Ohm	0.3	
MEASURE	OHMS LOWER	1500.00	1499.50	1500.50	1499.9	1499.9	Ohm	0.3	
MEASURE	OHMS LOWER	3200.00	3199.00	3201.00	3199.9	3199.9	Ohm	0.3	
		0	0	0					
MEASURE	OHMS RTD LOWER	350.00	349.80	350.20	350.02	350.02	Ohm	0.3	
		0	0	0					
MEASURE	TYPE K	-180.0	-180.6	-179.4	-180.20	-180.20	°C	1.2	
MEASURE	TYPE K	0.0	-0.6	0.6	-0.10	-0.10	°C	1.2	



Illiana Instrumentation Service LLC
 1831 Govert Drive
 Schererville, IN 46375



Cert #: 4102 Rev. 1 Item # 1468 Certificate Date: 9/1/2022

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT
MEASURE	TYPE K	400.0	399.4	400.6	399.80	399.80	°C	1.2	
MEASURE	TYPE K	800.0	799.2	800.8	799.80	799.80	°C	1.2	
MEASURE	TYPE K	1000.0	999.0	1001.0	999.80	999.80	°C	1.2	
MEASURE	TYPE K	1300.0	1298.7	1301.3	1299.80	1299.80	°C	1.2	
		0.0	0.0	0.0					
SOURCE	TYPE K	-180.00	-180.60	-179.40	-179.51	-179.51	°C	1.2	
SOURCE	TYPE K	0.00	-0.60	0.60	0.14	0.14	°C	1.2	
SOURCE	TYPE K	400.00	399.40	400.60	400.13	400.13	°C	1.2	
SOURCE	TYPE K	800.00	799.20	800.80	800.11	800.11	°C	1.2	
SOURCE	TYPE K	1000.00	999.00	1001.00	1000.14	1000.14	°C	1.2	
SOURCE	TYPE K	1300.00	1298.70	1301.30	1300.11	1300.11	°C	1.2	
		0.0	0.0	0.0					
MEASURE	TYPE J	-200.0	-200.6	-199.4	-200.3	-200.3	°C	1.2	
MEASURE	TYPE J	0.0	-0.6	0.6	-0.1	-0.1	°C	1.2	
MEASURE	TYPE J	300.0	299.4	300.6	299.8	299.8	°C	1.2	
MEASURE	TYPE J	600.0	599.4	600.6	599.8	599.8	°C	1.2	
MEASURE	TYPE J	900.0	899.1	900.9	899.8	899.8	°C	1.2	
MEASURE	TYPE J	1200.0	1198.8	1201.2	1199.8	1199.8	°C	1.2	
		0	0	0					
SOURCE	TYPE J	-200.00	-200.60	-199.40	-199.54	-199.54	°C	1.2	
SOURCE	TYPE J	0.00	-0.60	0.60	0.17	0.17	°C	1.2	
SOURCE	TYPE J	300.00	299.40	300.60	300.21	300.21	°C	1.2	
SOURCE	TYPE J	600.00	599.40	600.60	600.14	600.14	°C	1.2	
SOURCE	TYPE J	900.00	899.10	900.90	900.14	900.14	°C	1.2	
SOURCE	TYPE J	1200.00	1198.80	1201.10	1200.14	1200.14	°C	1.2	
		0.00000	0.00000	0.00000					
MEASURE	TYPE S	1000.0	999.0	1001.0	999.0	999.0	°C	1.2	
MEASURE	TYPE S	1150.0	1149.0	1151.0	1149.0	1149.0	°C	1.2	
MEASURE	TYPE S	1300.0	1299.0	1301.0	1299.0	1299.0	°C	1.2	
MEASURE	TYPE S	1450.0	1449.0	1451.0	1449.0	1449.0	°C	1.2	
MEASURE	TYPE S	1600.0	1599.0	1601.0	1599.0	1599.0	°C	1.2	
MEASURE	TYPE S	1750.0	1749.0	1751.0	1749.0	1749.0	°C	1.2	
		0.0	0.0	0.0					
SOURCE	TYPE S	1000.00	999.00	1001.00	999.84	999.84	°C	1.2	
SOURCE	TYPE S	1150.00	1148.90	1151.10	1149.83	1149.83	°C	1.2	
SOURCE	TYPE S	1300.00	1298.70	1301.30	1299.85	1299.85	°C	1.2	



Illiana Instrumentation Service LLC
 1831 Govert Drive
 Schererville, IN 46375



Cert #: 4102 Rev. 1 Item # 1468 Certificate Date: 9/1/2022

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT
SOURCE	TYPE S	1450.00	1448.60	1451.40	1449.88	1449.88	°C	1.2	
SOURCE	TYPE S	1600.00	1598.40	1601.60	1599.93	1599.93	°C	1.2	
SOURCE	TYPE S	1750.00	1748.30	1751.70	1749.89	1749.89	°C	1.2	
		0.0	0.0	0.0					
MEASURE	TYPE N	0.0	-0.6	0.6	-0.5	-0.5	°C	1.2	
MEASURE	TYPE N	200.0	199.4	200.6	199.6	199.6	°C	1.2	
MEASURE	TYPE N	500.0	499.4	500.6	499.7	499.7	°C	1.2	
MEASURE	TYPE N	800.0	799.2	800.8	799.7	799.7	°C	1.2	
MEASURE	TYPE N	1100.0	1098.9	1101.1	1099.7	1099.7	°C	1.2	
MEASURE	TYPE N	1300.0	1298.7	1301.3	1299.6	1299.6	°C	1.2	
		0.0	0.0	0.0					
SOURCE	TYPE N	0.00	-0.60	0.60	0.40	0.40	°C	1.2	
SOURCE	TYPE N	200.00	199.40	200.60	200.37	200.37	°C	1.2	
SOURCE	TYPE N	500.00	499.40	500.60	500.31	500.31	°C	1.2	
SOURCE	TYPE N	800.00	799.20	800.80	800.24	800.24	°C	1.2	
SOURCE	TYPE N	1100.00	1098.90	1101.10	1100.33	1100.33	°C	1.2	
SOURCE	TYPE N	1300.00	1298.70	1301.30	1300.29	1300.29	°C	1.2	
		0	0	0					
MEASURE	TYPE T	-200.0	-200.6	-199.4	-199.8	-199.8	°C	1.2	
MEASURE	TYPE T	0.0	-0.6	0.6	0.0	0.0	°C	1.2	
MEASURE	TYPE T	100.0	99.4	100.6	100.0	100.0	°C	1.2	
MEASURE	TYPE T	200.0	199.4	200.6	200.0	200.0	°C	1.2	
MEASURE	TYPE T	300.0	299.4	300.6	300.0	300.0	°C	1.2	
MEASURE	TYPE T	400.0	399.4	400.6	399.9	399.9	°C	1.2	
		0.0	0.0	0.0					
SOURCE	TYPE T	-200.00	-200.60	-199.40	-199.61	-199.61	°C	1.2	
SOURCE	TYPE T	0.00	-0.60	0.60	0.14	0.14	°C	1.2	
SOURCE	TYPE T	100.00	99.40	100.60	100.13	100.13	°C	1.2	
SOURCE	TYPE T	200.00	199.40	200.60	200.11	200.11	°C	1.2	
SOURCE	TYPE T	300.00	299.40	300.60	300.07	300.07	°C	1.2	
SOURCE	TYPE T	400.00	399.40	400.60	400.09	400.09	°C	1.2	
		0	0	0					
SOURCE	mADC LOWER	4.0000	3.9972	4.0028	4.0013	4.0013	maDC	0.01	
SOURCE	mADC LOWER	12.0000	11.9956	12.0044	12.0020	12.0020	maDC	0.018	
SOURCE	mADC LOWER	24.0000	23.9932	24.0068	24.0030	24.0030	maDC	0.018	
		0	0	0					



Cert #: 4102 Rev. 1 Item # 1468 Certificate Date: 9/1/2022

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT
SOURCE	mVDC LOWER	0.00000	-0.02000	0.02000	0.00130	0.00130	mVDC	0.0045	
SOURCE	mVDC LOWER	45.00000	44.97000	45.03000	45.01900	45.01900	mVDC	0.01	
SOURCE	mVDC LOWER	100.00000	99.96000	100.04000	100.02100	100.02100	mVDC	0.01	
		0	0	0					
SOURCE	VDC LOWER	0.000	-0.002	0.002	0.000000	0.000000	VDC	0.0000045	
SOURCE	VDC LOWER	5.000	4.997	5.003	4.999800	4.999800	VDC	0.00053	
SOURCE	VDC LOWER	10.000	9.996	10.004	9.999700	9.999700	VDC	0.0058	
		0.0	0.0	0.0					
SOURCE	OHMS LOWER	15.00	14.90	15.10	15.019	15.019	Ohm	0.0018	
SOURCE	OHMS LOWER	360.00	359.90	360.10	359.950	359.950	Ohm	0.016	
SOURCE	OHMS LOWER	500.00	499.50	500.50	500.000	500.000	Ohm	0.09	
SOURCE	OHMS LOWER	1500.00	1499.50	1500.50	1500.030	1500.030	Ohm	1.2	
SOURCE	OHMS LOWER	3200.00	3199.00	3201.00	3200.000	3200.000	Ohm	1.2	

Standards Used

I.D.	Model	Description	Manufacturer	Serial No.
1546	525B	TEMPERATURE/PRESSURE CAL	FLUKE	2041092
1731	34401A	DIGITAL MULTIMETER	HEWLETT PACKARD/AGILENT	3146A27542

Procedures Used

Name	Description
WI-523 CALIBRATORS	WI-523 CALIBRATORS