



Certificate of Calibration

Illiana Instrumentation Service LLC

1831 Govert Drive
Scherverville, IN 46375



Cert #: 20179 Rev. 1

Item # 1531

Certificate Date: 10/22/2024

CUSTOMER

ILLIANA INSTRUMENTATION SERVICE LLC
1831 GOVERT DRIVE
SCHERERVILLE, IN 46375

EQUIPMENT INFORMATION

Item # 1531
Manufacturer HONEYWELL
Model Number TVMUSX-888800-240-22-2-03
Serial Number 1126Y192349500001
Equip. Location MOBILE
Description RECORDER

CALIBRATION INFORMATION

Date received: 10/22/2024
Technician: PAUL GROLLA

Cal Date: 10/22/2024
Temperature: 73 F

Cal. Interval 3 MONTHS
Humidity 43
Tag 822582

Cal. Due Date: 1/14/2025
Cal location LAB

Department:
Notes:
Problems: None
Adjustments None
Equipment Notes

Sensitivity OKAY
Initial modification offset 0
Final modification offset 0
Note: If the modification offset is not zero, readings reflect zero
Limitations CALIBRATED ONLY ON CHANNELS THAT INCLUDE CAL DATA
Accuracy +/- 1 DEG F.



Illiana Instrumentation Service LLC

1831 Govert Drive
Scherverville, IN 46375



Cert #: 20179 Rev. 1

Item # 1531

Certificate Date: 10/22/2024

Quality and traceability statements

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for examination upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2. The certificate date on this report constitutes the issue date. No statement of conformity with specifications is made or implied on this certificate. However, measurement results are reviewed, where applicable, to establish where any measurement result exceeded the tolerance listed on this certificate. Measurement results exceeding the tolerance are indicated by an "F" in the Out Of Tolerance column. Add correction factor to as left to get corrected reading. Values are uncorrected for test instrument error.

Calibrated By:

Paul Grolla

Approved By:

Laura Grolla

Customer Approval

Calibration Data

Values found OOT and adjusted are indicated by "A" and items left OOT are indicated by "F"

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT	Corr. Factor
CH 1	TYPEK	0.0	-1.0	1.0	0.5	0.0	°F	1.2		0.0
CH 1	TYPEK	280.0	279.0	281.0	280.4	280.0	°F	1.2		0.0
CH 1	TYPEK	760.0	759.0	761.0	760.6	760.0	°F	1.2		0.0
CH 1	TYPEK	1240.0	1239.0	1241.0	1240.6	1240.0	°F	1.2		0.0
CH 1	TYPEK	1720.0	1719.0	1721.0	1720.8	1720.0	°F	1.2		0.0
CH 1	TYPEK	2200.0	2199.0	2201.0	2200.6	2200.0	°F	1.2		0.0
CH 2	TYPEK	0.0	-1.0	1.0	0.0	0.0	°F	1.2		0.0
CH 2	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 2	TYPEK	760.0	759.0	761.0	760.0	760.0	°F	1.2		0.0
CH 2	TYPEK	1240.0	1239.0	1241.0	1240.0	1240.0	°F	1.2		0.0
CH 2	TYPEK	1720.0	1719.0	1721.0	1720.1	1720.0	°F	1.2		0.0
CH 2	TYPEK	2200.0	2199.0	2201.0	2200.2	2200.0	°F	1.2		0.0
CH 3	TYPEK	0.0	-1.0	1.0	-0.2	0.0	°F	1.2		0.0



Illiana Instrumentation Service LLC

1831 Govert Drive
 Schererville, IN 46375



Cert #: 20179 Rev. 1

Item # 1531

Certificate Date: 10/22/2024

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT	Corr. Factor
CH 3	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 3	TYPEK	760.0	759.0	761.0	759.8	760.0	°F	1.2		0.0
CH 3	TYPEK	1240.0	1239.0	1241.0	1239.7	1240.0	°F	1.2		0.0
CH 3	TYPEK	1720.0	1719.0	1721.0	1719.8	1720.0	°F	1.2		0.0
CH 3	TYPEK	2200.0	2199.0	2201.0	2199.9	2200.0	°F	1.2		0.0
CH 4	TYPEK	0.0	-1.0	1.0	-0.1	0.0	°F	1.2		0.0
CH 4	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 4	TYPEK	760.0	759.0	761.0	759.6	760.0	°F	1.2		0.0
CH 4	TYPEK	1240.0	1239.0	1241.0	1240.0	1240.0	°F	1.2		0.0
CH 4	TYPEK	1720.0	1719.0	1721.0	1720.0	1720.0	°F	1.2		0.0
CH 4	TYPEK	2200.0	2199.0	2201.0	2200.0	2200.0	°F	1.2		0.0
CH 5	TYPEK	0.0	-1.0	1.0	0.3	0.0	°F	1.2		0.0
CH 5	TYPEK	280.0	279.0	281.0	280.2	280.0	°F	1.2		0.0
CH 5	TYPEK	760.0	759.0	761.0	760.2	760.0	°F	1.2		0.0
CH 5	TYPEK	1240.0	1239.0	1241.0	1240.4	1240.0	°F	1.2		0.0
CH 5	TYPEK	1720.0	1719.0	1721.0	1720.1	1720.0	°F	1.2		0.0
CH 5	TYPEK	2200.0	2199.0	2201.0	2200.4	2200.0	°F	1.2		0.0
CH 6	TYPEK	0.0	-1.0	1.0	0.0	0.0	°F	1.2		0.0
CH 6	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 6	TYPEK	760.0	759.0	761.0	760.0	760.0	°F	1.2		0.0
CH 6	TYPEK	1240.0	1239.0	1241.0	1240.0	1240.0	°F	1.2		0.0
CH 6	TYPEK	1720.0	1719.0	1721.0	1720.4	1720.0	°F	1.2		0.0
CH 6	TYPEK	2200.0	2199.0	2201.0	2200.0	2200.0	°F	1.2		0.0
CH 7	TYPEK	0.0	-1.0	1.0	0.0	0.0	°F	1.2		0.0



Illiana Instrumentation Service LLC

1831 Govert Drive
 Schererville, IN 46375



Cert #: 20179 Rev. 1

Item # 1531

Certificate Date: 10/22/2024

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT	Corr. Factor
CH 7	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 7	TYPEK	760.0	759.0	761.0	760.0	760.0	°F	1.2		0.0
CH 7	TYPEK	1240.0	1239.0	1241.0	1240.1	1240.0	°F	1.2		0.0
CH 7	TYPEK	1720.0	1719.0	1721.0	1720.2	1720.0	°F	1.2		0.0
CH 7	TYPEK	2200.0	2199.0	2201.0	2200.5	2200.0	°F	1.2		0.0
CH 8	TYPEK	0.0	-1.0	1.0	0.2	0.0	°F	1.2		0.0
CH 8	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 8	TYPEK	760.0	759.0	761.0	760.0	760.0	°F	1.2		0.0
CH 8	TYPEK	1240.0	1239.0	1241.0	1240.4	1240.0	°F	1.2		0.0
CH 8	TYPEK	1720.0	1719.0	1721.0	1720.2	1720.0	°F	1.2		0.0
CH 8	TYPEK	2200.0	2199.0	2201.0	2200.5	2200.0	°F	1.2		0.0
CH 9	TYPEK	0.0	-1.0	1.0	0.0	0.0	°F	1.2		0.0
CH 9	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 9	TYPEK	760.0	759.0	761.0	760.1	760.0	°F	1.2		0.0
CH 9	TYPEK	1240.0	1239.0	1241.0	1240.0	1240.0	°F	1.2		0.0
CH 9	TYPEK	1720.0	1719.0	1721.0	1719.9	1720.0	°F	1.2		0.0
CH 9	TYPEK	2200.0	2199.0	2201.0	2200.2	2200.0	°F	1.2		0.0
CH 10	TYPEK	0.0	-1.0	1.0	0.1	0.0	°F	1.2		0.0
CH 10	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 10	TYPEK	760.0	759.0	761.0	760.0	760.0	°F	1.2		0.0
CH 10	TYPEK	1240.0	1239.0	1241.0	1240.0	1240.0	°F	1.2		0.0
CH 10	TYPEK	1720.0	1719.0	1721.0	1720.4	1720.0	°F	1.2		0.0
CH 10	TYPEK	2200.0	2199.0	2201.0	2200.4	2200.0	°F	1.2		0.0
CH 11	TYPEK	0.0	-1.0	1.0	0.2	0.0	°F	1.2		0.0



Illiana Instrumentation Service LLC

1831 Govert Drive
 Schererville, IN 46375



Cert #: 20179 Rev. 1

Item # 1531

Certificate Date: 10/22/2024

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT	Corr. Factor
CH 11	TYPEK	280.0	279.0	281.0	280.0	280.0	°F	1.2		0.0
CH 11	TYPEK	760.0	759.0	761.0	760.1	760.0	°F	1.2		0.0
CH 11	TYPEK	1240.0	1239.0	1241.0	1240.1	1240.0	°F	1.2		0.0
CH 11	TYPEK	1720.0	1719.0	1721.0	1720.3	1720.0	°F	1.2		0.0
CH 11	TYPEK	2200.0	2199.0	2201.0	2200.1	2200.0	°F	1.2		0.0
CH 12	TYPEK	0.0	-1.0	1.0	0.0	0.0	°F	1.2		0.0
CH 12	TYPEK	280.0	279.0	281.0	279.9	280.0	°F	1.2		0.0
CH 12	TYPEK	760.0	759.0	761.0	760.0	760.0	°F	1.2		0.0
CH 12	TYPEK	1240.0	1239.0	1241.0	1240.2	1240.0	°F	1.2		0.0
CH 12	TYPEK	1720.0	1719.0	1721.0	1720.3	1720.0	°F	1.2		0.0
CH 12	TYPEK	2200.0	2199.0	2201.0	2200.1	2200.0	°F	1.2		0.0

Standards Used

I.D.	Model	Description	Manufacturer	Serial No.
1546	525B	TEMP/PRESSURE CALIBRATOR	FLUKE	2041092

Procedures Used

Name	Description
WI-527	CALIBRATION OF VIDEO RECORDERS