



Certificate of Calibration & Compliance For:
ILLIANA INSTRUMENTATION SERVICE
 1831 GOVERT DRIVE
 SCHERERVILLE, IN 46375



CALIBRATION
 NVLAP LAB CODE: 200496-0

GeoCorp, Inc.
 www.geocorpinc.com
 9010 River Road ▪ Huron, Ohio 44839
 P: 419-433-1101 F: 419-433-1102
 17025 Accredited Lab

1675

NVLAP is a U.S. Government entity
 administered by the NIST, an agency
 of the Department of Commerce.

Calibration #	Y-7291	Temperature	Reading		Correction		ASTM E29 AVG. CORRECTION
			FRONT	BACK	FRONT	BACK	
Degree: °	F	Point					
		100	99.9	99.9	0.1	0.1	0.1
P.O. #:	3946	200	200.1	200.1	-0.1	-0.1	-0.1
		400	399.5	399.4	0.5	0.6	0.6
Description:	Thermocouple Wire	600	599.8	599.7	0.2	0.3	0.2
TYPE	K	800	800.5	800.4	-0.5	-0.4	-0.4
500'	K20-2-HC	1000	1001.4	1001.2	-1.4	-1.2	-1.3
		1200	1202.2	1202.1	-2.2	-2.1	-2.2
		1400	1403.2	1403.1	-3.2	-3.1	-3.2
		1600	1603.3	1603.2	-3.3	-3.2	-3.2
		1800	1804.0	1803.9	-4.0	-3.9	-4.0
		2000	2004.8	2004.6	-4.8	-4.6	-4.7
		2200	2205.7	2205.5	-5.7	-5.5	-5.6
		2300	2306.1	2306.0	-6.1	-6.0	-6.0
Spool #	GK90116-8-7A						
Test Date:	3/12/2019						
Job #:	221364						

This Item meets ASTM E230 Special Limits Standard Limits New Calibration Recalibration

BAC 5621 L, Table V +/- 2 °F (+/- 1.1° C) to 1000° F, +/- 0.4% above AMS 2750 E Noble Metal +/- 1° F (+/- 0.6° C) or +/- 0.1%

BAC 5621 L, Table V, FL 4 +/- 2 °F (+/- 1.1° C) or +/- 0.4% AMS 2750 E Base Metal +/- 2° F (+/- 1.1° C) or +/- 0.4%

The calibrated equipment used for this test, listed below, is maintained in accordance with AMS 2750E, BAC 5621L (no PSD's), ANSI/NCSL Z540-3, ISO 10012, ISO/IEC 17025, RPS 953, PWA (F-40). Measurement results are traceable to standards maintained at the NIST. Test methods are per ASTM E220-13, ASTM E230/E230M-17, GEOCORP QW-001-O

NIST SENSOR: 292627 Calibrated: 12/21/2018 Due: 12/21/2019 Type: S	Primary Sensor: A-4119 Calibrated: 2/8/2019 Due: 2/8/2020 Type: S	Primary Inst: Keithley 2701 S/N: 1212892 Calibrated: 10/22/2018 Due: 10/22/2019	Humidity: 40% Nominal Ambient Temp: 72° F Immersion Depth: 24" tech: BW by: BW
--	---	---	--

Estimated uncertainty, k=2, 95%
 Thermocouple Types-K and N: -196° C to < 0° C = .71° C; 0° C to 1093 = .89° C; >1093 to 1344 = 1.9° C
 Thermocouple Type-J: -196° C to < 0° C = .70° C; 0° C to 760° C = .88° C

Noreen Roth
NOREEN ROTH
 APPROVED SIGNATORY

TYPE
K Thermocouple Wire

