



GeoCorp, Inc.

www.geocorpinc.com

9010 River Road • Huron, Ohio 44839

P: 419-433-1101 F: 419-433-1102

17025 Accredited Lab

Certificate of Calibration & Compliance For:
ILLIANA INSTRUMENTATION SERVICE
 1831 GOVERT DRIVE
 SCHERERVILLE, IN 46375



CALIBRATION

NVLAP LAB CODE: 200496-0

1678

NVLAP is a U.S. Government entity administered by the NIST, an agency of the Department of Commerce.

Calibration #	Y-7242	Temperature	Reading		Correction		ASTM E29 AVG. CORRECTION	
			FRONT	BACK	FRONT	BACK		
Degree:	° F	Point	32	32.2	32.3	-0.2	-0.3	-0.2
P.O. #:	3946	100	100.3	100.3	100.3	-0.3	-0.3	-0.3
		200	200.5	200.5	200.5	-0.5	-0.5	-0.5
Description:	Thermocouple Wire	400	400.8	400.8	400.8	-0.8	-0.8	-0.8
TYPE	K	600	601.0	601.1	601.1	-1.0	-1.1	-1.0
500'	K20-2-HG	800	801.4	801.4	801.4	-1.4	-1.4	-1.4
		1000	1002.0	1002.0	1002.0	-2.0	-2.0	-2.0
		1200	1202.6	1202.7	1202.7	-2.6	-2.7	-2.6
		1400	1403.0	1403.1	1403.1	-3.0	-3.1	-3.0
		1600	1603.4	1603.4	1603.4	-3.4	-3.4	-3.4
Spool #	GK90219-9-2							
Test Date:	3/12/2019							
Job #:	221364							

This Item meets ASTM E230 Special Limits Standard Limits New Calibration Recalibration
 BAC 5621 L, Table V, +/- 2 °F (+/- 1.1° C) to 1000° F, +/- 0.4% above AMS 2750 E Noble Metal +/- 1° F (+/- 0.6° C) or +/- 0.1%
 BAC 5621 L, Table V, FL 4 +/- 2 °F (+/- 1.1° C) or +/- 0.4% AMS 2750 E Base Metal +/- 2° F (+/- 1.1° C) or +/- 0.4%

The calibrated equipment used for this test, listed below, is maintained in accordance with AMS 2750E, BAC 5621L (no PSD's), ANSI/NC SL Z540-3, ISO 10012, ISO/IEC 17025, RPS 953, PWA (F-40). Measurement results are traceable to standards maintained at the NIST. Test methods are per ASTM E220-13, ASTM E230/E230M-17, GEOCORP QW-001-O

NIST SENSOR: 292627 Calibrated: 12/21/2018 Due: 12/21/2019 Type: S	Primary Sensor: A-4106 Calibrated: 2/5/2019 Due: 2/5/2020 Type: S	Primary Inst: Keithley 2701 S/N: 4096981 Calibrated: 10/23/2018 Due: 10/23/2019	Humidity: 40% Nominal Ambient Temp: 72° F Immersion Depth: 24" tech: AG by: AG
---	--	--	---

Estimated uncertainty, k=2, 95%
 Thermocouple Types-K and N: -196° C to < 0° C = .71° C; 0° C to 1093 = .89° C; >1093 to 1344 = 1.9° C
 Thermocouple Type-J: -196° C to < 0° C = .70° C; 0° C to 760° C = .88° C

Noreen Roth
NOREEN ROTH
 APPROVED SIGNATORY

TYPE **K Thermocouple Wire**

