

Certificate #: 1754064
 Purchase Order #: 4520

Calibration Performed By:
 PROBATA CORPORATION
 14600 Metro Plaza Blvd.
 PO BOX 4090 (73083)
 Edmond, OK 73013
 (405) 607-4813 / FAX (405) 607-4832

For:
 ILLIANA INSTRUMENTATION SERVICE
 1831 GOVERT DRIVE
 SCHERERVILLE IN,46375
 onsite:

Equipment Information

Manufacturer: FLUKE	Asset Number: 1693
Model: 700P04	Serial Number: 95050406
Description: PRESSURE MODULE	Department:
Range/Grad: 0 - 15 PSID / 0.001 PSI	Location:
Temp./RH: 68.0 degF, / 50.0 % rH	User:
Cal Date: 6/26/2024	Cal Procedure: FLUK-700 SERIES
Cal. Due Date: 6/26/2026	Revision Date: 11/30/2001

Calibration Results

Received Condition: MEASURED VALUES < LIMITS
Returned Condition: AS FOUND/AS LEFT DATA IS THE SAME
Notes:

Probata Calibration Standards Used for Certification

Asset #	Manufacturer	Model #	Description	Last Cal.	Cal. Due Date
0737	ADDITEL	ADT161-01-GP36	DIGITAL PRESSURE MODULE	4/25/2024	4/30/2025

Probata certifies that the above instrument meets or exceeds specifications as stated in the referenced procedure (unless otherwise noted). It has been calibrated using measurement standards traceable to the International System of Units (SI), through the National Institute of Standards and Technology (NIST) or other National Metrology Institutes participating in the CIPM-MRA or radiometric techniques or natural physical constants. This calibration complies with ANSI/NCSL Z540-1-1994, ISO 17025 and meets the quality standards set forth in ISO 9001. This document applies only to the calibration of the item described.

No statement of conformity with specifications is made or implied on this certificate. However, measurement results are reviewed, where applicable, to establish where any measurement result exceeded the manufacturer's specifications.

ALL RESULTS MAY BE INVALID BASED ON CONDITIONS AS RECEIVED. THIS CERTIFICATE MAY NOT BE REPRODUCED, EXCEPT IN FULL, WITHOUT THE WRITTEN APPROVAL OF PROBATA CORPORATION



Approved By: Jon Scruggs
JON SCRUGGS
 QUALITY SUPERVISOR

Calibrated By: Tanner Chambers
TANNER CHAMBERS
 QUALITY MANAGER

Calibration Results

Parameter Tested	Nominal	As Found	As Left	Low Limit	High Limit	EU (k=2)
Pressure Verification						
	3.000 psi	3.000		2.992	3.008	0.0011 psi
	6.000 psi	5.999		5.992	6.008	0.0011 psi
	9.000 psi	8.999		8.992	9.008	0.0011 psi
	12.000 psi	11.999		11.992	12.008	0.0011 psi
	15.000 psi	14.998		14.992	15.008	0.0011 psi

Range: 0.0 to 15.0 psi

Accuracy: 1 Year 0 to 50 °C ±0.05% of Span

Measurement results that are greater than limits of error are indicated by an "!"

The above measurement uncertainties represent expanded uncertainties expressed at approximately the 95% confidence level using a coverage factor of k=2

FRM-1601B rev. 11-17