

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Illiana Instrumentation
1831 Govert Drive
Schererville, IN 46375-

Date received: 03/21/21
Calibration date: 03/21/21
Recalibration due: 06/21/21

Purchase order:
Location of calibration: Lab

Item number: 1707	Item location: Lab
Tag number:	Temp: 72.0° F. Humidity: 45 Sensitivity: OK
Manufacturer: Honeywell, Inc.	Condition as found: In tolerance
Model num: TVMUSX-888000-100-22-0	Condition as left: In tolerance
Description: 1707 Recorder	Intentional offset as found: 0.0000
Serial number: 1121Y191869500001	Intentional offset as left: 0.0000
Accuracy: +/- 1 Deg F.	Limitations: None
Item notes:	

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the laboratory may cause the calibration of the above item to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for review upon request and approval of our Quality Manager. The uncertainties represent an expanded uncertainty expressed at @95% confidence level using a coverage factor of k=2. The date this report is signed constitutes the issue date. Pass/Fail criteria does not take into account measurement uncertainty.

Tech: *Cody Nimon* Date: 03/30/21 Approved by: *Laura Grolla* Date: 03/30/21
Cody Nimon - Technician Laura Grolla - Deputy Quality Manager

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	-0.4000	0.0000	1.0000	-1.0000	Deg F	CH 01 Type T	Yes

Standard(s): 1546 Fluke 525B
Adjustments: Adjusted bias and sensor comp values
Problems:
Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	49.6000	50.0000	51.0000	49.0000	Deg F	CH 01 Type T	Yes

Standard(s): 1546 Fluke 525B
Adjustments: Adjusted bias and sensor comp values
Problems:
Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-50.5000	-50.0000	-49.0000	-51.0000	Deg F	CH 01 Type T	Yes

Standard(s): 1546 Fluke 525B
Adjustments: Adjusted bias and sensor comp values
Problems:
Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	99.6000	100.0000	101.0000	99.0000	Deg F	CH 01 Type T	Yes

Standard(s): 1546 Fluke 525B
Adjustments: Adjusted bias and sensor comp values
Problems:
Uncertainty: 1.2000 Deg F Procedure: WI-508



ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.5000	-100.0000	-99.0000	-101.0000	Deg F	CH 01 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.0000	-150.0000	-149.0000	-151.0000	Deg F	CH 01 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	0.0000	0.0000	1.0000	-1.0000	Deg F	CH 02 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	49.6000	50.0000	51.0000	49.0000	Deg F	CH 02 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.9000	-50.0000	-49.0000	-51.0000	Deg F	CH 02 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	99.4000	100.0000	101.0000	99.0000	Deg F	CH 02 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.7000	-100.0000	-99.0000	-101.0000	Deg F	CH 02 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508



ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.6000	-150.0000	-149.0000	-151.0000	Deg F	CH 02 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	1.0000	0.0000	1.0000	-1.0000	Deg F	CH 03 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	51.0000	50.0000	51.0000	49.0000	Deg F	CH 03 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.0000	-50.0000	-49.0000	-51.0000	Deg F	CH 03 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	100.7000	100.0000	101.0000	99.0000	Deg F	CH 03 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.0000	-100.0000	-99.0000	-101.0000	Deg F	CH 03 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.0000	-150.0000	-149.0000	-151.0000	Deg F	CH 03 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	0.5000	0.0000	1.0000	-1.0000	Deg F	CH 04 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	50.1000	50.0000	51.0000	49.0000	Deg F	CH 04 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.4000	-50.0000	-49.0000	-51.0000	Deg F	CH 04 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	99.7000	100.0000	101.0000	99.0000	Deg F	CH 04 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.2000	-100.0000	-99.0000	-101.0000	Deg F	CH 04 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.9000	-150.0000	-149.0000	-151.0000	Deg F	CH 04 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	0.8000	0.0000	1.0000	-1.0000	Deg F	CH 05 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508



ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	50.8000	50.0000	51.0000	49.0000	Deg F	CH 05 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.3000	-50.0000	-49.0000	-51.0000	Deg F	CH 05 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	100.8000	100.0000	101.0000	99.0000	Deg F	CH 05 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.1000	-100.0000	-99.0000	-101.0000	Deg F	CH 05 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.0000	-150.0000	-149.0000	-151.0000	Deg F	CH 05 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	0.2000	0.0000	1.0000	-1.0000	Deg F	CH 06 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	49.8000	50.0000	51.0000	49.0000	Deg F	CH 06 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.6000	-50.0000	-49.0000	-51.0000	Deg F	CH 06 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	99.4000	100.0000	101.0000	99.0000	Deg F	CH 06 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.6000	-100.0000	-99.0000	-101.0000	Deg F	CH 06 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.6000	-150.0000	-149.0000	-151.0000	Deg F	CH 06 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	-0.3000	0.0000	1.0000	-1.0000	Deg F	CH 07 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	49.7000	50.0000	51.0000	49.0000	Deg F	CH 07 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-50.6000	-50.0000	-49.0000	-51.0000	Deg F	CH 07 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508



ILLIANA INSTRUMENTATION
1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com
CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	99.8000	100.0000	101.0000	99.0000	Deg F	CH 07 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-100.5000	-100.0000	-99.0000	-101.0000	Deg F	CH 07 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-150.6000	-150.0000	-14.9000	-151.0000	Deg F	CH 07 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	0.4000	0.0000	1.0000	-1.0000	Deg F	CH 08 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	50.4000	50.0000	51.0000	49.0000	Deg F	CH 08 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.6000	-50.0000	-49.0000	-51.0000	Deg F	CH 08 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	100.4000	100.0000	101.0000	99.0000	Deg F	CH 08 Type T	Yes

Standard(s): 1546 Fluke 525B
 Adjustments: Adjusted bias and sensor comp values
 Problems:
 Uncertainty: 1.2000 Deg F Procedure: WI-508



ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.4000	-100.0000	-99.0000	-101.0000	Deg F	CH 08 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.3000	-150.0000	-149.0000	-151.0000	Deg F	CH 08 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	-0.1000	0.0000	1.0000	-1.0000	Deg F	CH 09 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	50.0000	50.0000	51.0000	49.0000	Deg F	CH 09 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-49.8000	-50.0000	-49.0000	-51.0000	Deg F	CH 09 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	100.2000	100.0000	101.0000	99.0000	Deg F	CH 09 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.3000	-100.0000	-99.0000	-101.0000	Deg F	CH 09 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.4000	-150.0000	-149.0000	-151.0000	Deg F	CH 09 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
0.0000	0.0000	0.0000	1.0000	-1.0000	Deg F	CH 10 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
50.0000	50.1000	50.0000	51.0000	49.0000	Deg F	CH 10 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-50.0000	-50.0000	-50.0000	-49.0000	-51.0000	Deg F	CH 10 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	100.1000	100.0000	101.0000	99.0000	Deg F	CH 10 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-100.0000	-99.8000	-100.0000	-99.0000	-101.0000	Deg F	CH 10 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-150.0000	-149.8000	-150.0000	-149.0000	-151.0000	Deg F	CH 10 Type T	Yes

Standard(s): 1546 Fluke 525B

Adjustments: Adjusted bias and sensor comp values

Problems:

Uncertainty: 1.2000 Deg F

Procedure: WI-508

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Customer approval: _____ Date: _____

Total pages for Certificate 999999-01707-5474: 10