



ITEM 1711


**AMETEK Brookfield**  
**INSTRUMENTATION & SPECIALTY CONTROLS DIVISION**  
 11 Commerce Blvd. | Middleboro, MA 02346  
 P: 508.946.6200 | F: 508.946.6262  
 www.brookfieldengineering.com

**CERTIFICATE OF VISCOSITY STANDARD CALIBRATION**

<i>Viscosity Standard: 5000</i>		<i>Lot Number: 042021</i>	
<i>Certification/Issue Date: 04/20/21</i>		<i>Expiry Date: November 18, 2022</i>	
<b>Viscosity</b>		<b>Temperature</b>	
(centipoise)		(°C)	(°F)
4778		25.0	77.00
<b>Model</b>	<b>TEST INSTRUMENT</b>		<b>Serial Number</b>
LVDVNX			86017552
<b>Model</b>	<b>TEMPERATURE SENSOR</b>		<b>Serial Number</b>
1502A			C03315

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.  
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE



David J. Moonay, Ph.D -Quality Engineer - Rheologist