



Certificate of Calibration & Compliance For:
ILLIANA INSTRUMENTATION SERVICE
 1831 GOVERT DRIVE
 SCHERERVILLE, IN 46375

ITEM 1712



CALIBRATION
 NVLAP LAB CODE: 200496-0

GeoCorp, Inc.

www.geocorpinc.com

9010 River Road ▪ Huron, Ohio 44839
 P: 419-433-1101 F: 419-433-1102
 17025 Accredited Lab

NVLAP is a U.S. Government entity
 administered by the NIST, an agency
 of the Department of Commerce.

Calibration #	BB3422	Temperature	Reading		Correction		ASTM E29 AVG. CORRECTION
			FRONT	BACK	FRONT	BACK	
Degree:	° F	Point	FRONT	BACK	FRONT	BACK	
		32	32.1	32.1	-0.1	-0.1	-0.1
P.O. #:	4194	100	100.1	100.2	-0.1	-0.2	-0.2
		200	200.3	200.4	-0.3	-0.4	-0.4
Description:	Thermocouple Wire	400	400.8	400.9	-0.8	-0.9	-0.8
TYPE	K	600	601.0	601.1	-1.0	-1.1	-1.0
500'	K20-2-HG	800	801.4	801.5	-1.4	-1.5	-1.4
		1000	1001.8	1001.9	-1.8	-1.9	-1.8
		1200	1202.4	1202.5	-2.4	-2.5	-2.4
		1400	1403.1	1403.2	-3.1	-3.2	-3.2
		1600	1603.6	1603.7	-3.6	-3.7	-3.6
Spool Length:	500'						
Spool #	GI10423-6-27						
Test Date:	5/16/2021						
Job #:	249195						

This Item meets ASTM E230 Special Limits Standard Limits New Calibration Recalibration
 BAC 5621 L, Table V +/- 2 °F (+/- 1.1° C) to 1000° F, +/- 0.4% above AMS 2750 F Noble Metal +/- 1° F (+/- 0.6 ° C) or +/- 0.1%
 BAC 5621 L, Table V, FL 4 +/- 2 °F (+/- 1.1° C) or +/- 0.4% AMS 2750 F Base Metal +/- 2° F (+/- 1.1 ° C) or +/- 0.4%

The calibrated equipment used for this test, listed below, is maintained in accordance with AMS 2750F, BAC 5621L (no PSD's), ANSI/NCSS Z540-3, ISO 10012, ISO/IEC 17025, RRP 54000 , PWA (F-40). Measurement results are traceable to standards maintained at the NIST. Test methods are per ASTM E220-19, ASTM E230/E230M-17, GEOCORP QW-001-O

NIST SENSOR: 0000001650 Calibrated: 10/2/2020 Due: 10/2/2021 Type: S	Primary Sensor: A-4320 Calibrated: 2/24/2021 Due: 2/24/2022 Type: S	Primary Inst: Keithley 2701 S/N: 4096981 Calibrated: 7/8/2020 Due: 7/8/2021	Humidity: 40% Nominal Ambient Temp 72° F Immersion Depth: 24" tech: AG by: AG Issue Date: 5/16/2021
---	--	--	---

Estimated uncertainty, k=2, 95 %
 Thermocouple Types-K and N: -196 °C to < 0 °C = 0.71 °C; 0 °C to 1093 = 0.89 °C; >1093 to 1344 = 1.9 °C
 Thermocouple Type-J: -196 °C to < 0 °C = 0.70 °C; 0 °C to 760 °C = 0.88 °C
Noreen Roth
NOREEN ROTH
 APPROVED SIGNATORY

