



Calibration complies with ISO/IEC
17025, ANSI/NCSL Z540-1, and 9001

1762

Cert. No.: 1042-14565714

Traceable® Certificate of Calibration for Waterproof/Shockproof Stopwatch

Manufactured for and distributed by : Fisher Scientific "300 Industry Drive,,Pittsburgh,PA,15275-1001"

Instrument Identification:

Model: 06-662-56,11755833

S/N: 230726198

Manufacturer: Control Company

Standards/Equipment:

Description	Serial Number	Due Date	NIST Traceable Reference
Non-Contact Frequency Counter	26.66887	10 Jul 2024	1000493638

Certificate Information:

Technician: 447 Procedure: CAL-01 Cal Date: 31 Oct 2023 Cal Due Date: 31 Oct 2025
 Test Conditions: 41.58%RH 23.03°C 1025mBar

Calibration Data: (New Instrument)

Unit(s)	Nominal	As Found	In Tol	Nominal	As Left	In Tol	Min	Max	±U	TUR
sec/24hr	N.A.	N.A.		0.000	0.500	Y	-8.64	8.64	0.041	>4:1

This certificate indicates Traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

A Test Uncertainty Ratio of at least 4:1 is maintained unless otherwise stated and is calculated using the expanded measurement uncertainty. Uncertainty evaluation includes the instrument under test and is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement": (GUM). The uncertainty represents an expanded uncertainty using a coverage factor k=2 to approximate a 95% confidence level. In tolerance conditions are based on test results falling within specified limits with no reduction by the uncertainty of the measurement. The results contained herein relate only to the item calibrated. This certificate shall not be reproduced except in full, without written approval of Control Company.

Nominal=Standard's Reading; As Left=Instrument's Reading; In Tol=In Tolerance; Min/Max=Acceptance Range; ± U=Expanded Measurement Uncertainty; TUR=Test Uncertainty Ratio; Accuracy=t(Max-Min)/2; Min=As Left Nominal(Rounded) - Tolerance; Max= As Left Nominal(Rounded) + Tolerance;

Nicol Rodriguez

Nicol Rodriguez, Quality Manager

Note :

Maintaining Accuracy:

In our opinion once calibrated your Waterproof/Shockproof Stopwatch should maintain its accuracy. There is no exact way to determine how long calibration will be maintained. Waterproof/Shockproof Stopwatch change little, if any at all, but can be affected by aging, temperature, shock, and contamination.

Recalibration:

For factory calibration and re-certification traceable to National Institute of Standards and Technology contact Control Company.

Issue Date : 31 Oct 2023

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
 Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO/IEC 17025:2017 Calibration Laboratory Accredited by (A2LA) American Association for Laboratory Accreditation, Certificate No. 1750.01.
 Control Company is ISO 9001:2015 Quality Certified by DNV GL, Certificate No. CERT-01805-2006-AQ-HOU-ANAB.
 International Laboratory Accreditation Cooperation - Multilateral Recognition Arrangement (ILAC-MRA).