



CERTIFICATE OF ANALYSIS
Complies with ISO 17034, ISO Guide 31,
ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



This certificate indicates traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

1764

Certificate No. : 4889-14618824
Description : Buffer 10.012 pH Blue
Catalog Number : 15-078-287,15568703 **Lot :** CC791200
Certificate Date : 15 Nov 2023 **Expiration Date :** 15 Nov 2025
Certified Value : 10.004 pH **U:** ±0.011 pH (k=2) at 25°C

Certification measurements are performed under ISO 17034, A2LA accreditation no. 1750.02. They are traceable to recognized national and international standards via an unbroken chain of comparisons. pH is defined as the negative logarithm of the hydrogen ion activity.

MEASUREMENT: Minimum twelve (12) 100 ml samples were measured from this lot. The pH of each sample was determined using a pH meter and electrode.

UNCERTAINTY: The certified value is given as the average of the measured samples. The reported expanded uncertainty (U) is determined from the measurement variation from sample to sample, change due to shelf life, repeated use, and uncertainty of the measurement process. The value of uncertainty is multiplied by k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. Uncertainty is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement".

METHOD: The certified reference material is produced and analyzed by Control Company. The certified reference material is comprised of deionized water, 0.025 mol/kg H₂O sodium hydrogen carbonate, and 0.025 mol/kg H₂O sodium carbonate. Issue Date : 15 Nov 2023

Nicol Rodriguez
Nicol Rodriguez, Quality Manager

Traceability: Standards and Equipment Used

Description	Serial Number	Due Date	Traceable Reference
pH/Ion Meter	658R067-N023		
pH Reference Material 10.000	E1406422	03 Apr 2025	064-22
Digital Thermometer	102008443	17 Jul 2024	4000-14313687
pH Electrode	23151-F12		

Laboratory Environment Conditions: 39.00%RH 24.80°C 1021mBar

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO 17034:2016 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2015 (Reference Materials – Contents of Certificates and Labels), ISO 17034:2016 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2017 "Certification of Reference Materials – General and Statistical Principles". Control Company is an ISO/IEC 17025:2017 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2015 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-ANAB). Traceable® is a registered trademark of Control 3 Inc.



CERTIFICATE OF ANALYSIS
 Complies with ISO 17034, ISO Guide 31,
 ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



Temperature Correction Information:

Use the chart below only for making absolute measurements. That is, measurements without any automatic temperature correction (temperature coefficient set to 0).

Temperature Correction Chart												
°C	0.0	0.1	0.2	0.3	0.4	0.5	0.6	0.7	0.8	0.9		
18	10.076	10.075	10.074	10.073	10.072	10.071	10.069	10.068	10.067	10.066		
19	10.065	10.064	10.063	10.062	10.061	10.060	10.059	10.058	10.057	10.055		
20	10.054	10.053	10.052	10.051	10.050	10.049	10.048	10.047	10.046	10.045		
21	10.044	10.043	10.042	10.041	10.040	10.039	10.038	10.037	10.036	10.035		
22	10.034	10.033	10.032	10.031	10.030	10.029	10.028	10.027	10.026	10.025		
23	10.024	10.023	10.022	10.021	10.020	10.019	10.018	10.017	10.016	10.015		
24	10.014	10.013	10.012	10.011	10.010	10.009	10.008	10.007	10.006	10.005		
25	10.004	10.003	10.002	10.001	10.000	9.999	9.998	9.997	9.996	9.995		
26	9.994	9.993	9.992	9.991	9.990	9.989	9.988	9.987	9.986			
27	9.985	9.984	9.983	9.982	9.981	9.980	9.979	9.978	9.977			
28	9.976	9.975	9.974	9.973	9.972	9.971	9.970	9.969	9.968			

The above data are derived values based upon IUPAC data (Pure and Applied Chemistry 74, 2169-2200) and data/algorithm obtained using a temperature controlled calibration bath.

Shown on the chart is temperature (in the far left column) in whole degrees. Shown across the top row is temperature in tenths of a degree. Using a thermometer, measure the temperature of this certified reference material. Locate the measured temperature in whole numbers on the far left column. Follow across the row to the temperature in tenths of a degree. At the intersection is the certified reference material value at that specific temperature. Standardize the meter using that value.

Example: Measured temperature is 24.5°C. Find 24°C in the far left column, find the row 0.5°C. Where 24°C and 0.5°C intersect, read the value in pH.

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO 17034:2016 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (AZLA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2015 (Reference Materials - Contents of Certificates and Labels), ISO 17034:2016 Quality Management System Guidelines for the Production of Reference Materials, and ISO Guide 35:2010 Certification Requirements for General Services Providers. Control Company is also ISO 9001:2015 certified by American Association for Laboratory Accreditation (AZLA Certificate No. 1750.01). Control Company is ISO 9001:2015 certified by DNV GL (Certificate No. CER1-01805-2006-AQ-HOU-ANAB). Traceable® is a registered trademark of Control 3 Inc.