

1772



AMETEK Brookfield  
 INSTRUMENTATION & SPECIALTY CONTROLS DIVISION  
 11 Commerce Blvd. | Middleboro, MA 02346  
 P: 508.946.6200 | F: 508.946.6262  
 www.brookfieldengineering.com

**CERTIFICATE OF VISCOSITY STANDARD CALIBRATION**

Viscosity Standard: 1000		Lot Number: 101624	
Certification/Issue Date: 10/16/24		Expiry Date: April 24, 2026	
<b>Viscosity</b>		<b>Temperature</b>	
(centipoise)		(°C)	(°F)
984		25.0	77.00
<b>Model</b>	<b>TEST INSTRUMENT</b>	<b>Serial Number</b>	
LVDVNX		86017552	
<b>Model</b>	<b>TEMPERATURE SENSOR</b>	<b>Serial Number</b>	
1502A		B76401	

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.  
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist