



1773

AMETEK Brookfield  
 INSTRUMENTATION & SPECIALTY CONTROLS DIVISION  
 11 Commerce Blvd. | Middleboro, MA 02346  
 P: 508.946.6200 | F: 508.946.6262  
 www.brookfieldengineering.com

**CERTIFICATE OF VISCOSITY STANDARD CALIBRATION**

AMETEK BROOKFIELD VISCOSITY REFERENCE STANDARD		
Viscosity Standard: 5000		Lot Number: 042224
Certification/Issue Date: 04/22/24		Expiry Date: April 28, 2026
Viscosity	Temperature	
(centipoise)	(°C)	(°F)
4824	25.0	77.00
Model	TEST INSTRUMENT	Serial Number
LVDVNX		86055517
Model	TEMPERATURE SENSOR	Serial Number
1502A		B86724

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD  
 IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.  
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN  
 CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist